



Cross-Compatibility and United States Compliance

FERNO®

The following matrix provides information regarding **cross-compatibility of cot, cot fastener, and patient restraint systems**. It is meant as a quick reference guide for emergency services that use both Ferno and Stryker products. Where applicable, the combinations below have been dynamically crash tested to the three United States industry recognized consensus standards.

Compatible Systems

STANDARD	=	COT			+	COT FASTENER		+	PATIENT RESTRAINT			
		FERNO POWER X1 Ambulance Cot	Stryker Performance- PRO XT	Stryker Power-PRO XT	FERNO PRO F1 Universal Cot Fastener	Stryker Performance- LOAD	FERNO Rail/Antler Series	Stryker Rail/Antler Series*	FERNO X-Restraints	FERNO BioSafe Restraints	Stryker X-Restraints	Stryker Rugged-X Restraints*
KKK-A-1822F through change notice 11 (SAE J3027)		•			•				•			
		•			•					•		
		•				•			•			
			•		•						•	
				•	•						•	
			•		•							•
				•	•							•
CAAS-GVS v1.0 (2016) (SAE J3027)		•			•				•			
		•			•					•		
		•				•			•			
			•		•						•	
				•	•						•	
			•		•							•
				•	•							•
NFPA 1917 (2016) (SAE J3027)		•			•				•			
		•			•					•		
		•				•			•			
			•		•						•	
				•	•						•	
			•		•							•
				•	•							•
KKK-A-1822F through change notice 8 (AMD 004)		•					•		•			
		•					•		•	•		
		•						•	•			
		•						•		•		
			•				•	•			•	•
				•			•	•			•	•
							•	•				•

*As advertised

When used in conjunction with the manufacturer supplied **floor plate system** and **safety hook**, each of the combinations listed in the matrix comply with the respective standard listed.